



# COMPRESSED NITROUS OXIDE

Medical Grade

## COMPOSITION

Nitrous Oxide Gas

## SPECIFICATION

Min 98%

## IMPURITIES

MOISTURE

< 67 ppm

NO/NO<sub>2</sub>

< 2 ppm

A certificate of conformity is issued with every cylinder.  
Testing is done on batches basis.

General Information	
<b>Valve Outlet</b>	BS 341
<b>Cylinder size</b>	No-14
<b>Cylinder Material</b>	J
<b>Weight Kg</b>	MS
	30

Pakistan Oxygen Limited (Previously Linde Pakistan Limited)  
PO Box # 4845, Dockyard Road, West Wharf Karachi-74000,  
Pakistan.  
[www.pakxygen.com](http://www.pakxygen.com)

